

I currently use a glass filter why should I use the HSQ Kit solutions instead for qualification – FAQ

Glass filters for UV detectors are not considered primary standards whereas the solutions in the MicroSolv HSQ Kit $^{\text{m}}$  are primary standards.

Click HERE for PQ and HSQ ordering information and pictures.

Printed from the Chrom Resource Center
Copyright 2024, All Rights Apply
MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com