

What is the mobile phase used in the HSQ Kit and how much is used – FAQ

For isocratic system testing, prepare one liter of 13:87 acetonitrile:0.1% acetic acid (1mL/L).

For gradient testing, prepare 2 liters.

Click [HERE](#) for PQ Kit™ and HSQ Kit™ ordering information and pictures.