



## Do the different tests in the HSQ Kit for HPLC use the same method conditions - FAQ

**Yes, this is why it can be run overnight from an autosampler.**

Simply set your conditions, program the system for performance qualification (PQ) validation and let it run. Use the following conditions:

Wavelength:	273nm
Injection Volume:	10µl (except for volume <b>linearity</b> )
Flow Rate:	1 mL/minute
Temperature:	Ambient