



Are column oven temperatures qualified by the HSQ Kit for HPLC – FAQ

No, the use of column ovens are not required for PQ testing of your HPLC instruments. Therefore, there is no temperature protocols included in the kit.

Click [HERE](#) for PQ Kit™ and HSQ Kit™ ordering information and pictures.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com