



## Do different retention times for peaks in the Resolution Test Mixture on different systems impact the integrity of my data - FAQ

### **No impact at all.**

There will certainly be mobile phase preparation differences, aging of columns, etc. The HSQ Kit™ instructions call for adjustment of the mobile phase to meet system suitability, which for example in our C of A, calls to adjust the mobile phase to achieve a 4-5 minute retention time window. Keep in mind that the C o A merely confirms that the column was tested, and met specs when it left our facility.

Click [HERE](#) for PQ and HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024