



Storing the HSQ & PQ Kit test solutions when not in use – How To

The test solutions should be stored at room temperature 20-25°C in a safe, dry place. *Never refrigerate these solutions.*

Click [HERE](#) for PQ & HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-05-2024