



Can I use the MicroSolv Window Maker for removal of PFA coating on copper wires – FAQ

The unit is designed specifically for use in making “detection windows” for polyimide coated fused quartz silica with a very specific thickness and outer diameter (OD) which are used for capillary electrophoresis instruments.

The low amount of heat that is delivered from the heating unit is specific and designed to burn off this polyimide and not damage the underlying quartz or silica. The unit is sold and supported only for the above single use.

However, many fiber optics companies have purchased the unit and are using it successfully with glass wires and cables; but not all are successful. If you wish to “burn off” the PFA on wire, it is possible that unit would be sufficient but the amount of PFA removed each time may be inconsistent. Therefore we do not recommend our unit for this application

Click [HERE](#) for *MicroSolv Window Maker ordering information and pictures.*

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 10-11-2024