

## Determine dwell volume of gradient HPLC systems – How To

**With some instruments the gradient dwell volume can be quite significant, even a few milliliters.**

At the start of the gradient, the volume in the instrument is filled with the mobile phase corresponding to the initial gradient conditions and consequently the “dwell volume” of the mobile phase should flow through the column before the starting gradient profile arrives at the top of the column. As a result the expected gradient **elution** is delayed and some analytes, especially weakly retained ones, may migrate a certain distance along the column during this unintended initial isocratic step, which contributes in this way to the **elution** volume.

This behavior can be described as a two-step **elution** with the first, isocratic (*dwell volume*) step, followed by the second, gradient, step. The dwell volume may differ from one instrument to another and may cause difficulties if a gradient HPLC method is transferred from one instrument to another one. To avoid these problems and to make possible precise prediction of the gradient **elution** data by calculation, the gradient dwell volume should be accounted for in method development and appropriate correction should be adopted for the instrumental gradient delay in calculations.

### 1. Theory

The major application of HPLC using gradient conditions is for reversed-phase (*RP-HPLC*) systems<sup>1</sup>, however the general conclusions are also applicable to other HPLC modes (*normal-phase*, **Aqueous Normal Phase ANP**, *ion-exchange*, etc.). For RP-HPLC separations, the solute retention factor  $k$  can be approximately related to the volume-fraction of organic modifier ( $B$ ) in a mobile phase  $\varphi$  as:

where  $k_w$  is the (*extrapolated*) value of  $k$  for water as the mobile phase ( $\varphi=0$ ), and  $S$  is a constant characteristic for each solute in the sample. Retention time in a linear-gradient separation can be expressed as:

$$t_R = (t_0/b) \log(2.3k_0b + 1) + t_0 + t_d \quad (2)$$

Here,  $t_0$  is the column dead-time,  $k_0$  is the value of  $k$  at the start of the gradient, and  $b$  is a gradient steepness parameter given by equation 3 below:

$$b = t_0 \Delta\varphi S / t_G = V_m \Delta\varphi S / t_G F = (\Delta\varphi / t_G) (V_m S / F) \quad (3)$$

where  $\Delta\varphi$  is the change in the volume-fraction of organic modifier ( $B$ ) during the gradient ( $\Delta\varphi=0.01\%B$ ),  $V_m$  is the column dead-volume, and  $F$  is the flow-rate. The quantity  $t_d$  is the equipment hold-up or ‘dwell’ time, equal to the hold-up or ‘dwell’ volume  $V_0$  divided by flow-rate. The equipment dwell volume is often wrongly assumed to be negligible

( $V_D=0$ ), which leads to the problems. Relative retention  $t_R/t_0$  can be defined, and from Eq. 2 (assuming  $t_d=0$ ):

$$t_R/t_0 = (1/b) \log(2.3k_0b + 1) + 1 \quad (4)$$

If column dimensions or flow-rate are changed,  $t_0=V_m/F$  will change. However, as long as  $Q$  is held constant (by varying  $t_G$ ),  $b$  and  $t_R/t_0$  remain constant for each solute, and relative retention will not change. Eq. 2 and Eq. 4 are actually only approximations when  $V_D>0$ , but it is still possible to calculate exact values of  $t_R$  as a function of  $k_0$ ,  $t_0$ ,  $b$  and  $V_D$ . Resolution in gradient **elution** can be expressed in terms that are similar to isocratic separation:

$$R_s = (1/4)(\alpha - 1)N^{1/2}(k^* / [k^* + 1]) \quad (5)$$

where  $k^*$  (the average retention factor  $k$  during gradient **elution**) is given as:

$$k^* = 1/[1.15b + (1/k_0)] \quad (6)$$

For peaks that elute later in the gradient, and for which  $k_0>1.15b$ ,

$$k^* = 1/1.5b \quad (6a)$$

## 2. Dwell-volume effects

According to Eq. 4 above, relative retention should remain unchanged as long as  $b$  (and  $Q$ ) are constant, assumes that  $V_D=0$ . When there is a significant dwell volume  $V_D$  in the instrument, this situation changes. As mentioned in the opening paragraph gradient conditions are delayed by a dwell time  $t_D=V_D/F$ . As a result analytes is under isocratic conditions (for a time  $t_D$ ) with  $k=k_0$ . As a result the value of  $b$ , the gradient steepness parameter ( $b=0$  for isocratic **elution**) is smaller than given by 3 for the gradient **elution**. The composite value of  $b$  ( $b'$ ; i.e. effective gradient steepness) for a separation of a sample undergoing isocratic **elution** at the start is therefore smaller. The effective gradient steepness  $b'$  is therefore a combination of initial **elution** with  $b = 0$  and **elution** with  $b$  calculated by equation 3. As a result solutes which are weakly retained (elute early) will undergo the most migration during isocratic part of the run and as a consequence for these compounds  $b'$  will be greatest.

For later eluting compounds  $b'/b$  will approach 1 and become smaller as  $t_R$  becomes smaller. This combined gradient steepness parameter  $b'$  will depend both on  $k_0$  and the ratio of dwell volume to column volume ( $V_D/V_m$ ). To assure there are minimal changes in relative retention of the compounds if a gradient HPLC method is transferred from one instrument to another, one should keep both  $Q$  and ( $V_D/V_m$ ) constant. If  $V_D/V_m$  is held constant while changing column dimensions, relative retention ( $t_R/t_0$ ) should remain unchanged.

<sup>1</sup> J. W. Dolan and L. R. Snyder; ***“Maintaining fixed band spacing when changing column dimensions in gradient elution”*** J. Chromatogr. A, Vol. 799, 13 March 1998, Pages 21-34.



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-20-2024