
What is the UV wavelength that I should use for CELixir-SDS™? - INTERNAL

INTERNAL ONLY - THIS ITEM IS DISCONTINUED

If a Diode Array detector is available, it is suggested to use it. However, any individual wavelength should be selected depending on the UV maxima of your samples. There are no restriction on the UV range of CELixir-SDS™.

[Click here for more information about CELixir-SDS](#)