



## What is the difference between a union and an adapter in HPLC - FAQ

The biggest difference is the price where a union is much less expensive than an adapter.

For performance considerations, a union will make a connection between tubing and fittings that have the same thread size and thread count configurations. An adapter will connect two different thread sizes and thread count.

Click [HERE](#) for HPLC fittings ordering information and pictures.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024