

If I do not remove gas from my mobile phase solvents what are the consequences - FAQ

HPLC solvents that are exposed to the atmosphere, even for a short time have the ability to absorb compressible carbon dioxide CO₂ and oxygen O₂.

When your HPLC is operating with low back pressure, the actual linear flow of materials could be different then when you have higher back pressure. This is due to the compressibility of air versus the compressibility of gas. This means that your HPLC results could different retention times and different peak shapes.

Also, noisy baselines occur which can lead to very unreliable results in both quantitation and retention times. Some columns, most notably polymeric columns, do not tolerate outgassing very well. It is therefore always recommended to keep your solvents optimally performing by degassing them and keeping them degassed.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-20-2024