



What is the pH stability of Cogent HPS HPLC columns – FAQ

These columns are extremely stable in from pH 2.0 to pH 8.6.

Depending on the concentration of the acid or base will determine if these columns would be stable outside of this range.

[Click here to view more information about Cogent HPS HPLC columns](#)



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-12-2024