



## What is the pH stability of Cogent HPS HPLC columns - FAQ

**These columns are extremely stable in from pH 2.0 to pH 8.6.**

Depending on the concentration of the acid or base will determine if these columns would be stable outside of this range.

[Click here to view more information about Cogent HPS HPLC columns](#)

