
Are there any solvents that I cannot use with the knitted reactor coils (KRC) and knitted delay tubes (KDT)?

Yes, you cannot use any perfluorinated solvents or any solvent not compatible with PTFE with the KRC and KDT.

[Click here for ordering information for the post column reactor](#)