

The only restriction on whether or not you can use the knitted reactor coils KRC is the inlet tubing of your detector. The absolute minimum ID can be 0.008". It is preferred that the tubing ID be 0.01". Anything smaller than 0.008" is not compatible.

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com