



Can I change the filter element in the Cogent Column Filters - FAQ

No, there is no need to replace the stainless steel filter element (frit) in the pre-column, Cogent Column Filters. When the unit is ready to be replaced, simply dispose of it. The device will be destroyed if you open it or try to remove the filter element.

[Click here to view more information about pre-column filters for HPLC](#)



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-18-2024