

**No, there is no need to replace the stainless steel filter element (frit) in the pre-column, Cogent Column Filters.** When the unit is ready to be replaced, simply dispose of it. The device will be destroyed if you open it or try to remove the filter element.

*[Click here to view more information about pre-column filters for HPLC](#)*



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)