



## Can I change the filter element in the Cogent Column Filters - FAQ

**No, there is no need to replace the stainless steel filter element (frit) in the pre-column, Cogent Column Filters.** When the unit is ready to be replaced, simply dispose of it. The device will be destroyed if you open it or try to remove the filter element.

*[Click here to view more information about pre-column filters for HPLC](#)*

