



Which solvents should not be used with the Cogent Column Filters – FAQ

All solvents that are NOT compatible with PEEK should not be used with the pre-column, Cogent Column Filters. They include but not limited to: DMSO, Methylene Chloride and concentrated THF.

[Click here to view more information about pre-column filters for HPLC](#)



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-12-2024