



## Which solvents should not be used with the Cogent Column Filters – FAQ

**All solvents that are NOT compatible with PEEK should not be used with the pre-column, Cogent Column Filters. They include but not limited to: DMSO, Methylene Chloride and concentrated THF.**

*[Click here to view more information about pre-column filters for HPLC](#)*

