



## Which solvents should I avoid with AQ syringe filters that have a nylon membrane – FAQ

**Aldehydes and strong acids such as hydrochloric and nitric should be avoided with AQ™ nylon syringe filters or membrane filters.** Click [HERE](#) for syringe filters ordering information and pictures.

### Attachment

**MicroSolv filters equivalency study pdf** [Download File](#)

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-03-2024