

Aldehydes and strong acids such as hydrochloric and nitric should be avoided with AQ™ nylon syringe filters or membrane filters. Click [HERE](#) for syringe filters ordering information and pictures.

Attachment

MicroSolv filters equivalency study pdf [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com