
Which solvents should I avoid with AQ syringe filters that have a nylon membrane - FAQ

Aldehydes and strong acids such as hydrochloric and nitric should be avoided with AQ™ nylon syringe filters or membrane filters. Click [HERE](#) for syringe filters ordering information and pictures.

Attachment

MicroSolv filters equivalency study pdf [Download File](#)