

Do you retain samples from each lot?

All MicroSolv syringe filters have production lot numbers which can be followed to complete traceability to the raw materials used in manufacturing the filters. We have a mandatory 3 year archive time of samples and lot analysis.

Click [HERE](#) for AQ syringe filters ordering information and pictures.

Attachment

MicroSolv filter equivalency study pdf [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com