



Are the MicroSolv AQ brand of syringe filter membranes supported by extra membranes in the plastic housing – FAQ

The PTFE Teflon syringe filter membranes are backed with an inert polypropylene membrane for support.

All the others syringe filters including Nylon, Regenerated Cellulose, Cellulose Acetate, PVDF etal do not require additional support and therefore have no additional membranes.

Click [HERE](#) for ordering information and pictures of AQ brand syringe filters

Attachment

MicroSolv Filters Equivalency Study pdf [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 09-26-2024