

By suppressing their ionization at low pH, acids will be more like neutral compounds and will retain on Reversed Phase HPLC columns much the same way that neutral compounds will. The same is true for using bases with basic compounds.

Another benefit of running on Type A and Type B silica at low pH is that the silanol activity is reduced. This should sharpen your peaks.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com