

**Suggestion:** Try substituting formic acid in DI water and adding ammonium acetate (*10mM*) and acetic acid (*1%*) to the aqueous phase. This should improve the reproducibility of peak areas and retention times.

**In addition,** extensive reconditioning of the analytical column with pure acetonitrile in between runs is essential for good reproducibility when working with plasma samples.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)