

This is a question that is always asked and the following are only guidelines. For exact details, consult the FDA, your regulatory department or the USP.

Change of Detector Wavelength	Not Allowed (you can vary up to 3nm in <b>accuracy</b> but you cannot adjust the detector.)
Column ID	+/- 50%
Column Length	+/- 70%
Column Temperature	+/- 20%
Concentration of <b>buffer</b> salts	+/- 10%
Flow Rate	+/- 50%
Particle Size	- 50%
pH of Mobile Phase	+/-0.2 units
Ratio of Mobile Phase Solvents	+/-10% absolute (This adjustment is not suggested)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)