



Is The Vial Centrifuge NRTL rated – FAQ

No, not at this time. The Vial Centrifuge™ is not rated under the Nationally Recognized Testing Laboratories (NRTL) however, it is CE marked which indicates that the product complies with European health, safety, and environmental production standards.

Click [HERE](#) for ordering information.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024