





## Alprazolam Analyzed with HPLC - AppNote

### **Robust Separation of API from USP Internal Standard**

The USP Assay Method for Alprazolam uses a bare Silica Column and a complex Mobile Phase consisting of Acetonitrile, Chloroform, Butyl Alcohol, and Acetic Acid. In this Method, a simple LCMS compatible Mobile Phase is used and produces excellent Peak shapes for both the API and its USP internal standard.

Furthermore, a resolution of 4.3 was obtained between the two Peaks, which meets the USP system suitability of  $R_s \geq 2.0$ . Two impurity Peaks are also observed, which further illustrates the resolution capabilities of the analysis. This Method illustrates how the MS-compatible HPLC-UV Methods described in various application notes can be successfully adapted for LCMS.





Peak 1

Peak 2 **Peaks:**

1. Triazolam (internal standard)
2. Alprazolam (API)
- 3, 4. Impurities

### Method Conditions

**Column:** Cogent Diamond Hydride™, 4µm, 100Å

**Catalog No.:** 70000-7.5P

**Dimensions:** 4.6 x 75mm

#### Mobile Phase:

A: DI Water / 0.1% Formic Acid (v/v)

B: Acetonitrile / 0.1% Formic Acid (v/v)

#### Gradient:

Time (minutes)

%B



0	95
1	95
6	50
7	95

**Post Time:** 3 minutes

**Injection vol.:** 1 $\mu$ L

**Flow rate:** 1.0mL / minute

**Detection:** UV @ 254nm

**Sample Preparation:**

*Tablet:* A 0.25mg strength generic Xanax® tablet was ground and added to a 10mL volumetric flask. After diluting with 50% Solvent A / 50% Solvent B, it was sonicated for 10 minutes. A portion was filtered with a 0.45 $\mu$ m Nylon Syringe Filter (MicroSolv Tech Corp.).

*Internal Standard:* 1mg / mL Triazolam in Methanol diluent.

*Working Solutions:* 20 $\mu$ L of the internal standard and 980 $\mu$ L of the tablet extract were mixed. Peak identities were confirmed by individual solutions of the tablet extract and the internal standard.

$t_0$ : 0.9 minutes

**Note:** Alprazolam is a member of the benzo-diazepine class of compounds, prescribed to treat a variety of anxiety-related conditions.

**MICROSOLV** TECHNOLOGY CORPORATION  
HPLC Columns and Chromatography Accessories



## Attachment

**No 183 Alprazolam Xanax Analyzed with HPLC pdf** 0.6 Mb [Download File](#)