


Alprazolam Analyzed with HPLC - AppNote

Robust Separation of API from USP Internal Standard

The USP Assay Method for Alprazolam uses a bare Silica Column and a complex Mobile Phase consisting of Acetonitrile, Chloroform, Butyl Alcohol, and Acetic Acid. In this Method, a simple LCMS compatible Mobile Phase is used and produces excellent Peak shapes for both the API and its USP internal standard.

Furthermore, a resolution of 4.3 was obtained between the two Peaks, which meets the USP system suitability of Rs≥2.0. Two impurity Peaks are also observed, which further illustrates the resolution capabilities of the analysis. This Method illustrates how the MS-compatible HPLC-UV Methods described in various application notes can be successfully adapted for LCMS.



1. Triazolam (internal standard)



2. Alprazolam (API)

3, 4. Impurities

Method Conditions

Column: Cogent Diamond Hydride[™], 4μm, 100Å

Catalog No.: 70000-7.5P **Dimensions:** 4.6 x 75mm

Mobile Phase:

A: DI Water / 0.1% Formic Acid (v/v) B: Acetonitrile / 0.1% Formic Acid (v/v)

Gradient:

| Time (minutes) | %B |
|----------------|----|
| 0 | 95 |
| 1 | 95 |
| 6 | 50 |
| 7 | 95 |

Post Time: 3 minutes Injection vol.: 1µL

Flow rate: 1.0mL / minute Detection: UV @ 254nm **Sample Preparation:**

Tablet: A 0.25mg strength generic Xanax® tablet was ground and added to a 10mL volumetric flask. After diluting with 50% Solvent A / 50% Solvent B, it was sonicated for 10 minutes. A portion was filtered with a 0.45µm Nylon Syringe Filter (MicroSolv Tech Corp.).

Internal Standard: 1mg / mL Triazolam in Methanol diluent.

Working Solutions: 20µL of the internal standard and 980µL of the tablet extract were mixed. Peak identities were confirmed by individual solutions of the tablet extract and the internal standard.

t₀: 0.9 minutes

Note: Alprazolam is a member of the benzo-diazepine class of compounds, prescribed to treat a variety of anxietyrelated conditions.



Attachment

No 183 Alprazolam Xanax Analyzed with HPLC pdf 0.6 Mb Download File MicroSolv Technology Corporation 9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com