


Low Molecular Weight Organic Acids

Low mass Acids were retained and Separated in this Method with the Aqueous Normal Phase mode of HPLC. The Method was able to resolve the three compounds in 10 minutes. Limits of Detection (LOD) for them ranged from 0.05 to 24 micromolar.



Peaks:

- 1. Maleic Acid 115.0031 m/z (M-H)-
- 2. Trans-Aconitic Acid 173.0086 m/z (M-H)-
 - 3. Citric Acid 191.0192 m/z (M-H)-

Method Conditions

Column: Cogent Diamond Hydride™, 4µm, 100Å

Catalog No.: 70000-15P-2 **Dimensions:** 2.1 x 150mm

Mobile Phase:

A: DI Water / 0.1% Ammonium Acetate
B: Acetonitrile / 0.1% Ammonium Acetate

Gradient:

Time (minutes)	%B
0	90
2	90
5	70
6	70
6.1	30
7	30
7.1	30

Printed from the Chrom Resource Center Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com



10

90

Flow rate: 400µL / minute

Detection: ESI - neg - Agilent 6210 MSD TOF Mass Spectrometer

Sample Preparation: Sample mixture was prepared in 50:50 DI Water / Acetonitrile

Notes: ANP Chromatography with MS Detection (ANP-LCMS) has become more and more popular for analysis of certain mixtures of organic acids because of the Simplicity, Rapidity and Stability of the Method. The statutory Methods for the determination of low molecular mass organic acids are Turbidimetric and Colorimetric Methods and Ion-Exchange Chromatography. "Turbidimetry by means of calcium oxalate precipitation and colorimetry by means of Zinc Ferricyanide color development are time-consuming and only achieve limited rather than exact results".



Attachment

No 64 Citric, Aconitic & Maleic Acids Analyzed with LCMS pdf 0.2 Mb Download File

Printed from the Chrom Resource Center Copyright 2024, All Rights Apply **MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

> Email: customers@mtc-usa.com Website: www.mtc-usa.com