



Serum Corticosterones Concentration Determinations

Serum Corticosterone Concentrations are of clinical significance in Adrenal dysfunction. Its measurement is sometimes used to diagnose apparent mineral Corticoid Excess Syndrome. It can also be used as a bio marker of malignancy in Adrenal Tumors.

Using this Method allows the simultaneous measurement of the two main Estrogen fractions, Estrone and Estradiol in Breast Tumor tissue. Highly sensitive assays for measuring the compounds at low levels of Estrogen in postmenopausal women, and monitoring Estrogen levels in women receiving hormone replacement therapy, can also be developed using the method presented here.



Peaks:

Parent Ion /mz	Printed from the Chrom Resource Center
1. Andrenosterone 301 283 (MH+ - H2O)	Copyright 2024, All Rights Apply
2. Corticosterone 347 329 (MH+ - H2O)	MicroSolv Technology Corporation
3. 4-Androstene-3,17-dione 287 319 (MH+ - O2)	9158 Industrial Blvd. NE, Leland, NC 28451
4. 11-Alpha-Acetoxyprogesterone 373 313(MH+ -H2GHOOH)	tel. (732) 380-8900, fax (910) 769-9435
5. Estrone 271 253 (MH+ -H2O)	Email: customers@mtc-usa.com

6. Estradiol 273 254 (MH+ -H₂O)**Method Conditions:****Column:** Cogent UDC Cholesterol™, 4µm, 100Å**Catalog No.:** 69069-7.5P**Dimensions:** 4.6 x 75mm**Mobile Phase:** 50:50 DI Water / Methanol / 0.5% Formic Acid**Injection vol.:** 5µL**Flow rate:** 0.5ml / minute**Detection:** APCI, Single Ion Monitoring Mass Spectrometer

Notes: The total ion chromatogram of a mixture of six Steroids presented, shows the resolving power of the Cogent UDC Cholesterol Column. The peak shapes are excellent (As < 1.1 for all solutes) using an isocratic Mobile Phase.

**Attachment****No 35 Corticosterones Analyzed with LCMS pdf** 0.2 Mb [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.comWebsite: www.mtc-usa.com