

Are nylon membrane and syringe filters compatible with chloroform – FAQ

All nylon filters are not recommended for use with halogenated solvents such as chloroform or with strong acids (1N and stronger).

For these liquids, you should consider using PTFE filters.

Click [HERE](#) for ordering information and pictures of AQ brand of syringe filters.