
Autoclaving Inserts with Plastic Springs is Not Recommended.

Autoclaving glass or plastic Inserts with Attached Plastic Springs is not recommended.

While the glass can withstand temperatures up to 1,000°F, the Polyethylene Spring will become deformed, may crack and will lose its functionality.

You can Sterilize these Inserts through Ethylene Oxide Gas Sterilization.



[Inserts Product Page](#)