

**The maximum operating temperature for the Cogent Diamond Hydride™ HPLC column is 60° C.**

However, Higher Temperature Columns are available on request with High Temperature all Stainless Steel Hardware.



*[Click HERE for Cogent Diamond Hydride Ordering Information](#)*

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)