



What is the maximum temperature for Cogent Diamond Hydride HPLC columns – FAQ

The maximum operating temperature for the Cogent Diamond Hydride™ HPLC column is 60° C.

However, Higher Temperature Columns are available on request with High Temperature all Stainless Steel Hardware.



[Click HERE for Cogent Diamond Hydride Ordering Information](#)

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com