



What is the recommended pH range for Cogent Diamond Hydride HPLC columns - FAQ

The recommended pH range for the Cogent Diamond Hydride™ HPLC column is 2.5 - 7.0



[Click HERE for Cogent Diamond Hydride Column Ordering Information](#)

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 08-07-2024