



What is the recommended pH range for Cogent Diamond Hydride HPLC columns - FAQ

The recommended pH range for the Cogent Diamond Hydride™ HPLC column is 2.5 - 7.0



[Click HERE for Cogent Diamond Hydride Column Ordering Information](#)