

## Can I sterilize screw caps for autosampler vials that have silicone rubber with PTFE septa - FAQ

Yes, if you use heat / steam or gamma irradiation, the AQ™, MicroSolv™ and the BASIK™ brands of 8mm, 9mm and 10mm screw caps can be sterilized or autoclaved for up to 15 minutes.

Click [HERE](#) for ordering information and pictures of autosampler vials