


Isocratic Method for Prednisone Tablets

This Method uses the Mobile Phase specified in the USP Method for Assay of Prednisone tablets. The resolution between the API and the Acetanilide Internal Standard meets the system suitability.

Resolution was found to be superior compared to standard C18 HPLC Columns. In addition, this Method also produces excellent precision under these conditions, as shown in the figure overlay.



1. Acetanilide (Internal Standard)

2. Prednisone (API)

Method Conditions

Column: Cogent Bidentate C8™, 4µm, 100Å

Catalog No.: 40008-75P **Dimensions:** 4.6 x 75mm

Mobile Phase: 68.8% DI Water / 25.0% Tetrahydrofuran (THF) / 6.2% Methanol (v/v) the Chrom Resource Center

Injection vol.: 10μL Copyright 2024, All Rights Apply

Flow rate: 1.0mL / minute

MicroSolv Technology Corporation

Detection: UV @ 254nm 9158 Industrial Blvd. NE, Leland, NC 28451

Sample Preparation: 10mg strength Prednisone tablet was ground and added to (3.25) mg volume tric (1976). Howas 435 diluted to mark with Methanol and sonicated 10 minutes. It was then filtered with a 0.45 mm. Nylon Syrjage Filter.com

Website: www.mtc-usa.com



(MicroSolv Technology Corp.). $110\mu L$ of the filtrate and $90\mu L$ of a 0.1mg / mL Acetanilide solution were combined and diluted with $800\mu L$ Methanol. Peaks were confirmed by individual standards.

to: 0.8 minutes

Note: Prednisone is used to treat symptoms due to low corticosteroid levels. This encompasses a wide variety of applications such as for Arthritis, severe allergic reactions, Multiple Sclerosis, and Lupus. It is marketed under a variety of trade names, such as Deltasone®, Meticorten®, and Orasone®.



Attachment

No 189 Prednisone Analyzed with HPLC pdf 0.5 Mb Download File

Printed from the Chrom Resource Center Copyright 2024, All Rights Apply **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Website: www.mtc-usa.com

Email: customers@mtc-usa.com