



Can the Ultra-T brand of PEEK tubing be used in medical applications - FAQ

In general, PEEK is FDA approved as a long term implantent and can be sterilized. This refers only to the natural type of PEEK tubing and does not include tubing with stripes or markings. The dashed stripes of the Ultra-T™ tubing can be removed with acetone if desired.

However, these tubing items are designed for laboratory use and using PEEK tubing or any MicroSolv product is not intended by the company for use in human or animal medical devices and has no FDA approval for such use.

Click [HERE](#) for high pressure tubing ordering information and pictures.

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com