

Metal impurities testing via ICP (Inductively Coupled Plasma) analysis is performed for the following, along with acceptance specifications:

Al: Not more than 5.0 ppm

Fe: Not more than 20.0 ppm

Ti: Not more than 0.5 ppm

Mg: Not more than 5.0 ppm

Ca: Not more than 15.0 ppm

Na: Not more than 20.0 ppm

It is important to note that the actual levels may be significantly lower than those in the table.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com