

## What is the pressure capability of the ARE brand all PEEK column coupler – FAQ

These ARE-Applied Research brand HPLC column couplers are actually unions (*cat# 47450-01, 47450-02, 47450-03, or 47450-04*) and are standard HPLC items that if installed properly while using standard temperatures and normal HPLC solvents, they should be able to withstand pressures up to 6,000 psi.

Click [HERE](#) for column coupler ordering information and pictures.

