



A white film developed on my MS source so is this column packing material – FAQ

If the film was from dissolved stationary phase material, you would see a significant distortion in peak shape for all your analytes due to the void that would be developing in the column. If it was from the stationary phase being hydrolyzed, you would see a change in retention and/or peak shape for your compounds.

Be sure that you are operating within the specifications of the HPLC column. It is in fact possible to dissolve the packing material or hydrolyze the bonded phase if the column is used Improperly.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-05-2024