



Can I use the holmium oxide solution in a UV instrument that calls for Starna or NIST cuvettes - FAQ

If your instrument manual calls for the SRM 2034 holmium oxide HOX solution, ([PQA-HOX-01](#)) you can use our HOX solution instead. The SRM 2034 item has been discontinued by NIST but our HOX solution is fully NIST traceable to their published standard. Therefore our HOX solution will be entirely suitable for this purpose.

Click [HERE](#) for Holmium Oxide solution ordering information and pictures.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-20-2024