

The UV Calibration Kit test replacement absorbance solutions are made in controlled laboratory conditions of 20–25 °C.

Within this temperature range, the variability on CofA values is not significant and therefore not reported. The stray light test, which is temperature sensitive, includes a temperature correction factor in the software, allowing the user to correct for unusual laboratory temperatures. Testing is independent of other laboratory conditions such as relative humidity.

Click [HERE](#) for UV Calibration Kit ordering information and pictures.

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)