
Is stainless steel or titanium better for mobile phase inlet filters - FAQ

Stainless steel is suitable for most applications but titanium is recommended for analyses involving biological analytes and matrices.

[Stainless steel inlet filter product page](#)