



Is stainless steel or titanium better for mobile phase inlet filters - FAQ

Stainless steel is suitable for most applications but titanium is recommended for analyses involving biological analytes and matrices.

[Stainless steel inlet filter product page](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com