

Yes, the Chrom Syringes are fully compatible in an autoclaving application.

If you are using the chromatography syringes with syringe filters, please verify that the filter is also suitable for the autoclaving conditions as well.

[Luer-Lock Syringe Product Page](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com